

93-853

Durable Nitrile Exam Glove with Extended Cuff and Advanced Barrier Protection

Medical exam gloves with extended chemical splash protection, approved for handling chemotherapy drugs

- Strong formulation: MICROFLEX® 93-853 exam gloves' robust nitrile formulation makes them more durable and resistant to rips, snags and tears
- Specialized protection: Simulating real-world overdose conditions, these gloves are specially tested against gastric acid (vomit) and fentanyl, as well as cytotoxic/chemotherapy drugs*
- Real-world performance: These sturdy medical exam gloves also satisfy NFPA 1999 Emergency Medical Glove Requirements, making them suitable for healthcare professionals, first responders and lab workers
- Low AQL: These medical exam gloves' low 0.65 AQL reinforces their barrier integrity
- Extended protection: Featuring an extended cuff, these gloves offer forearm and wrist protection for medical workers



*see product packaging for details

Industries

- Automotive
- Life Sciences
- Healthcare
- Chemical
- Machinery and Equipment

Recommended For

- Inspection, selecting, checking parts
- Assembly and inspection of components
- Picking and fastening components
- Equipment repair and maintenance
- General Purpose Auto Aftermarket
- Tire rotation & changes
- Animal care feeding and cleanup
- Maintaining lab furniture and equipment
- Sample taking and processing
- Administering drugs
- Raw material sample collection



TECHNICAL DATA SHEET

| | Product Information |
|------------------------|--|
| Material | Nitrile |
| External Glove Surface | Textured Fingers |
| Audit Standards | ISO 13485, ISO 9001 |
| Product Certification | Medical Device Regulation (EU) 2017/745 Class I, Personal Protective Equipment Regulation (EU) 2016/425 Category III risks, CE 0493, EN 1149-1/2/3, EN 16523-1, EN ISO 374-1:2016 Type B KPT, EN ISO 374-5:2016 Virus, EN 455 1-4, EN ISO 21420:2020, ISO 11193, ISO 18889:2019, PMDA Notification Number: 13B3X10307AH0001, Trade Name: Ansell Non Latex Medical Examination Glove, UKCA,EU Food Contact Approved |
| Packaging Overview | 50 gloves per dispenser 10 dispensers per case 500 gloves per case *40 gloves per dispenser (Size XXXL) *400 gloves per case (Size XXXL) |
| Storage Instructions | Keep out of direct sunlight; store in a cool and dry place. Keep away from sources of ozone or ignition. |
| Country Of Origin | Malaysia |
| Product Segmentation | High Risk |
| Antistatic | Yes EN1149 |
| Silicone Free | No |

PHYSICAL PROPERTIES

| | Typical Values | | Testing Method |
|---|---------------------|-------------|---------------------|
| Length (mm/inches) | 295 / 11.6 | | ASTM D3767,EN 420 |
| Freedom from Holes (Inspection level I) | 0.65 AQL | | ASTM D5151,EN 455-1 |
| Palm Thickness (mm/mils) | 0.14 / 5.5 | | ASTM D3767,EN 420 |
| Finger Thickness (mm/mils) | 0.22 / 8.7 | | ASTM D3767,EN 420 |
| | BEFORE AGING | AFTER AGING | |
| Ultimate Tensile Strength (MPa) | 32 | 27 | ASTM D412 & D573 |
| Elongation at Break (%) | 550 | 500 | ASTM D412 |
| Force at break (N) | 15 | 12 | EN 455-2 |

ORDERING INFORMATION

| Size | S (6.5 - 7) |
|--------------|-------------|
| Product Code | 93853070 |

Performance Standards and Regulatory Compliance













For additional information visit us at www.ansell.com, or call us at

Europe, Middle East & Africa Region Ansell Healthcare Europe NV T: +32 (0) 2 528 74 00 F: +32 (0) 2 528 74 01

Asia Pacific Region MY +603 8310 6688 apac.medical@ansell.com CN: +86 21 38275000 JP: +813 5549 8151

North America Region Ansell Healthcare Products LLC T: +1 800 800 0444 F: +1 800 800 0445

Latin America & Caribbean Region BR +55 (11) 3356-3100 latam.medical@ansell.com MX: +52 442 296 2050 latam.medical@ansell.com CO: +57 1 288 3247 latam.medical@ansell.com

Ansell Limited T: +61 1800 337 041 F: +61 1800 803 578 Russia

Ansell PYC Ten. +7 495 258 13 16

Ansell, ® and ™ are trademarks owned by Ansell Limited or one of its affiliates. US Patented and US and non-US Patents Pending: www.ansell.com/patentmarking © 2025 Ansell Limited. All Rights Reserved.

Neither this document nor any other statement made herein by or on behalf of Ansell should be construed as a warranty of merchantability or that any Ansell product is fit for a particular purpose. Ansell assumes no responsibility for the suitability or adequacy of an end user's selection of gloves for a specific application.

